



STEP 1

Remove all straps and open device.



STEP 2

Place foot into cushion, ensuring heel is located correctly.



STEP 3

Secure straps. Position de-rotation wedge to control leg position if required. Ensure touch tape is fully engaged.



STEP 4

If plantar flexion control is required use optional elastic straps to pull foot/ankle into appropriate position.